



Council of the  
European Union

Brussels, 21 November 2018  
(OR. en)

14561/18  
ADD 1

ENV 798  
MI 871  
DELECT 154

#### COVER NOTE

---

From:	Secretary-General of the European Commission, signed by Mr Jordi AYET PUIGARNAU, Director
date of receipt:	19 November 2018
To:	Mr Jeppe TRANHOLM-MIKKELSEN, Secretary-General of the Council of the European Union

---

No. Cion doc.:	C(2018) 7505 final - Annex
Subject:	ANNEX to the Commission Delegated Directive amending, for the purposes of adapting to scientific and technical progress, Annex III to Directive 2011/65/EU of the European Parliament and of the Council as regards an exemption for lead as activator in the fluorescent powder of discharge lamps containing phosphors

---

Delegations will find attached document C(2018) 7505 final - Annex.

---

Encl.: C(2018) 7505 final - Annex



Brussels, 16.11.2018  
C(2018) 7505 final

ANNEX

ANNEX

to

**Commission Delegated Directive**

**amending, for the purposes of adapting to scientific and technical progress, Annex III to Directive 2011/65/EU of the European Parliament and of the Council as regards an exemption for lead as activator in the fluorescent powder of discharge lamps containing phosphors**

## ANNEX

In Annex III, entry 18(b) is replaced by the following:

"18(b)	Lead as activator in the fluorescent powder (1% lead by weight or less) of discharge lamps when used as sun tanning lamps containing phosphors such as BSP (BaSi <sub>2</sub> O <sub>5</sub> :Pb)	Expires on: <ul style="list-style-type: none"><li>– 21 July 2021 for categories 1-7 and 10;</li><li>– 21 July 2021 for categories 8 and 9 other than in vitro diagnostic medical devices and industrial monitoring and control instruments;</li><li>– 21 July 2023 for category 8 in vitro diagnostic medical devices;</li><li>– 21 July 2024 for category 9 industrial monitoring and control instruments, and for category 11.</li></ul>
18(b)-I	Lead as activator in the fluorescent powder (1 % lead by weight or less) of discharge lamps containing phosphors such as BSP (BaSi <sub>2</sub> O <sub>5</sub> :Pb) when used in medical phototherapy equipment	Applies to categories 5 and 8, excluding applications covered by entry 34 of Annex IV, and expires on 21 July 2021."